

**DEPARTMENT OF APPLIED PHYSICS
DELHI TECHNOLOGICAL UNIVERSITY
BAWANA ROAD, DELHI- 110042**

Date: 08/11/2023

NOTICE INVITING QUOTATIONS

Sealed Quotations on company's letter head with GSTIN no. are invited for the procurement of Vapors Closet system in CFMRL lab, Applied Physics Department with the descriptions mentioned below. The sealed quotations can be submitted in the HOD office of Department of Applied Physics, DTU by 14th November 2023.

| S.No | Description of Item | Make | Quantity |
|-------------|-------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|-------------|-----------------|
| 1. | Vapors Closet Specifications: Power coated MS / GI SS / Jet black granite Fluorescent tube light Centrifugal exhaust blower (direct drive motor) Water tap Sink with drainage system Electrical sockets, FRP coating on inner walls Air foil, Scrubber system Storage cabinet Caster wheels w/ brake Wet & dry service valves Certification (ASHRAE110- 1995 / EN-14175-2003) Air Flow Monitor Temperature indicator Standard Size/ Dimension: NL-422FH 4 x 2 x 2 1219 x 610 x 610 | Any | 1 no. |

Dr. Bharti Singh

Department of Applied Physics,

Delhi Technological University

Bawana Road, Delhi -110042